

Thoracolumbar Spine Protocol

The protocol below represents a full Thoracolumbar spine exam. There is significant variation depending on the clinical question and clinician preference. It is highly suggested to acquire at last one post contrast series with fat saturation if contrast is administered. It is also highly suggested to acquire one series in the dorsal plane, typically a T2 or STIR. For scanning technique, see the [Thoracolumbar Spine MRI Guide](#)

Disclaimer: Please keep in mind any parameters, positioning, or coil selections listed are meant to serve as a guideline, your particular scanner specifications and coil set may vary, necessitating different parameters than those listed.

Position: Supine

Coil: Flex coil, Quad Knee, Spine Coil, or Knee Coil

Parameters:

- FOV: Sm (18cm) Med (24cm) Lg (32cm)
- Resolution: 0.6mm x 0.8mm
- Slice Thickness: 2-3mm
- Averages/NEX: 4

Sagittal T2 Lumbar
Sagittal T2 Thoracic
Sagittal STIR ROI
Sagittal T1 ROI
Axial T2 ROI
Axial T1 ROI
Contrast
Sagittal T1 ROI
Axial T1 ROI

Optional Sequences

3D Myelogram
HASTE Myelogram
Multi-echo GRE

From:

<https://wiki.virtual-scan.com/> - Knowledge Portal

Permanent link:

https://wiki.virtual-scan.com/doku.php?id=library:thoracolumbar_spine_protocol&rev=1750862970

Last update: 2025/06/25 14:49

